

**INDEPENDENT TESTING LAB**

830 ROBINWOOD COURT, TRAVERSE CITY, MI 49686

PH: 231-929-0905

FAX: 231-929-0894

www.gtanalytical.com

Company: VILLAGE OF KALKASKA
Name:
ClientProj: WEEKLY
GTA ProjNo: 061521-3

Site Addr: 1005 ISLAND LAKE RD
KALKASKA
Sampled By: BRENDON WOLF/VOK
Date Rec: 6/15/2021
Time Rec: 11:30 AM

Sample No.	Sample ID	Date Sampled	Time Sampled	Sample Matrix
1	INFLUENT	6/15/2021	7:36 AM	WATER
2	EFFLUENT	6/15/2021	7:46 AM	WATER

ELECTRONIC SIGNATURE REPORT. This is a final report for the following pages of data for the samples specified above. All analysis was performed by the methods stated and all quality control measures required were completed. All quality control information is available upon request.

Kirk Chase

Kirk L. Chase/Chemist
Grand Traverse Analytical
830 Robinwood Court
Traverse City, MI 49686
Ph: 231-929-0905
Fx: 231-929-0894
SP: 231-590-0291
kirk@gtanalytical.com

**INDEPENDENT TESTING LAB**

830 ROBINWOOD COURT, TRAVERSE CITY, MI 49686

PH: 231-929-0905

FAX: 231-929-0894

www.gtanalytical.com

COMPANY: VILLAGE OF KALKASKA

PROJECT NO: WEEKLY

LOCATION: 1005 ISLAND LAKE RD
KALKASKA
MI

SAMPLED BY: BRENDON WOLF/VOK

SAMPLE ID: INFLUENT

GTA PROJECT NO: 061521-3

GTA SAMPLE NO: 1

DATE RECEIVED: 6/15/2021
TIME RECEIVED: 11:30 AMDATE SAMPLED: 6/15/2021
TIME SAMPLED: 7:36 AM

SAMPLE MATRIX: WATER

SAMPLE ANALYSIS

<u>Analysis</u>	<u>Concentration</u>	<u>LOD</u>	<u>Units</u>	<u>Analyst</u>	<u>Date Completed</u>
AMMONIA-N EPA 350.1	27	0.30	mg/L (PPM)	KC	6/15/2021

ND = NOT DETECTED, RESULT < LOD

LOD = LIMIT OF DETECTION

s.u. = STANDARD pH UNITS REPORTED AT 25 C

DISS = DISSOLVED

**INDEPENDENT TESTING LAB**

830 ROBINWOOD COURT, TRAVERSE CITY, MI 49686

PH: 231-929-0905

FAX: 231-929-0894

www.gtanalytical.com

COMPANY: VILLAGE OF KALKASKA
PROJECT NO: WEEKLY
LOCATION: 1005 ISLAND LAKE RD
KALKASKA
MI
SAMPLED BY: BRENDON WOLF/VOK
SAMPLE ID: EFFLUENT

GTA PROJECT NO: 061521-3
GTA SAMPLE NO: 2
DATE RECEIVED: 6/15/2021
TIME RECEIVED: 11:30 AM
DATE SAMPLED: 6/15/2021
TIME SAMPLED: 7:46 AM
SAMPLE MATRIX: WATER

SAMPLE ANALYSIS

<u>Analysis</u>	<u>Concentration</u>	<u>LOD</u>	<u>Units</u>	<u>Analyst</u>	<u>Date Completed</u>
AMMONIA-N EPA 350.1	<0.030	0.030	mg/L (PPM)	KC	6/15/2021
NITRATE-N EPA 300.0	0.76	0.10	mg/L (PPM)	LV	6/15/2021
NITRITE-N EPA 300.0	<0.10	0.10	mg/L (PPM)	LV	6/15/2021

ND = NOT DETECTED, RESULT < LOD

LOD = LIMIT OF DETECTION

s.u. = STANDARD pH UNITS REPORTED AT 25 C

DISS = DISSOLVED

1 QC EPA 300.0 ANIONS

CONTINUING CAL

DATE: 6/15/2021

	<u>TRUE PPM</u>	<u>OBS PPM</u>	<u>%DEV</u>	<u>+/-10%?</u>
NITRATE-N	3.52	3.42	1.4	PASS
NITRITE-N	3.01	2.97	0.7	PASS

BLANK

DATE: 6/15/2021

	<u>OBS PPM</u>	<u><0.1PPM?</u>
NITRATE-N	0	PASS
NITRITE-N	0	PASS

SPIKE/DUP SPIKE

DATE: 6/15/2021

SAMPLE: PROCEDURAL BLANK

	<u>SPK OBS</u>	<u>TRUE</u>	<u>%REC</u>	<u>80-120%?</u>
NITRATE-N	1.62	1.84	88.0	PASS
NITRITE-N	5.88	6.08	96.7	PASS

	<u>DSPK OBS</u>	<u>TRUE</u>	<u>%REC</u>	<u>80-120%?</u>
DUP SPK NOT DONE				
NITRATE-N	1.66	1.84	90.2	PASS
NITRITE-N	5.99	6.08	98.5	PASS

	<u>SPK</u>	<u>DSPK</u>	<u>%DEV</u>	<u>+/-10%?</u>
NITRATE-N	88.0	90.2	-1.2	PASS
NITRITE-N	96.7	98.5	-0.9	PASS

3 QC AMMONIA-N EPA 350.1

CONTINUING CAL

DATE: 6/15/2021

	<u>TRUE PPM</u>	<u>OBS PPM</u>	<u>%DEV</u>	<u>+/-10%?</u>
AMMONIA-N	0.50	0.47	3.1	PASS

PROCEDURAL BLANK

DATE: 6/15/2021

	<u>OBS PPM</u>	<u><0.01PPM?</u>
AMMONIA-N	0	PASS

SPIKE/DUP SPIKE

DATE: 6/15/2021

SAMPLE: 060121-2-2

	<u>SPK OBS</u>	<u>TRUE</u>	<u>%REC</u>	<u>80-120%?</u>
AMMONIA-N	0.49	0.5	98.0	PASS

	<u>DSPK OBS</u>	<u>TRUE</u>	<u>%REC</u>	<u>80-120%?</u>
AMMONIA-N	0.49	0.5	98.0	PASS

	<u>SPK</u>	<u>DSPK</u>	<u>%DEV</u>	<u>+/-15%?</u>
AMMONIA-N	98	98	0.0	PASS

830 Robinwood Court
Traverse City, MI 49686
Phone (231) 929-0905 Fax (231) 929-0894

[illegible]